

Reports, Prosthetics Research Board
1954

Olson

May 1954



INDEX 45
24,155

Artificial Limbs

A Review of
Current Developments

ADVISORY COMMITTEE on ARTIFICIAL LIMBS

National Academy of Sciences
National Research Council

ADVISORY COMMITTEE on ARTIFICIAL LIMBS
Division of Engineering and Industrial Research
NATIONAL ACADEMY OF SCIENCES—NATIONAL RESEARCH COUNCIL

Paul E. Klopsteg, *Chairman*

Rufus H. Alldredge

Howard D. Eberhart

Robert R. McMath

C. Leslie Mitchell

Craig L. Taylor

T. Campbell Thompson

Philip D. Wilson

PROFESSIONAL ASSOCIATES

Robert S. Allen

Verne T. Inman

EXECUTIVE DIRECTOR

F. S. Strong, Jr.

EXECUTIVE SECRETARY

A. Bennett Wilson, Jr.

Artificial Limbs

BRYSON FLEER, *Editor*

MAY 1954

A publication of the Advisory Committee on Artificial Limbs, National Academy of Sciences—National Research Council, issued three times a year, in January, May, and September, in partial fulfillment of Veterans Administration Contract VAm-21223. Copyright 1954 by the National Academy of Sciences—National Research Council. Quoting and reprinting are freely permitted, providing appropriate credit is given. The opinions expressed by contributors are their own and are not necessarily those of the Advisory Committee on Artificial Limbs.

CONTENTS

LIMBS IN LIMBO		
	Herbert Eftman	1
THE OBJECTIVES OF THE LOWER-EXTREMITY PROSTHETICS PROGRAM		
	Howard D. Eberhart	4
CONTRIBUTIONS OF THE LOWER-EXTREMITY PROSTHETICS PROGRAM		
	Edmond M. Wagner	8
MECHANICAL AIDS FOR ALIGNMENT OF LOWER-EXTREMITY PROSTHESES		
	Charles W. Radcliffe	20
STATUS OF THE ABOVE-KNEE SUCTION SOCKET IN THE UNITED STATES		
	Chester C. Haddan and Atha Thomas	29
DIGEST OF MAJOR ACTIVITIES OF THE ARTIFICIAL LIMB PROGRAM		40

ADVISORY COMMITTEE ON ARTIFICIAL LIMBS

NATIONAL ACADEMY OF SCIENCES—NATIONAL RESEARCH COUNCIL

2101 Constitution Ave.

Washington 25, D. C.



Limbs in Limbo

HERBERT ELFTMAN, Ph.D.¹

TO STAND on his feet and to walk with his legs wherever his heart desires are natural rights guaranteed to man by his own constitution. Heads may plan and hands may build, but only where legs and feet have brought them. Loss of the lower limb is therefore a major catastrophe.

When loss of leg occurs, replacement becomes the primary hope. Ages past, an unknown man hobbled forth from his cave in search of a willow; with one of its limbs adopted as his own, he walked back with majesty. Since then the stage of history has resounded with the staccato echo of countless amputees marching with peg-leg, grit, and gumption.

Rapid perfection of limb construction was to be anticipated after these early ventures had focused human ingenuity upon the problem. To the superlative talent which mankind has shown in the production of machinery both intricate and sturdy, the building of a mechanical leg would appear to offer little difficulty. Why is it, then, that artificial limbs have so generally belonged to the limbo of things undeserving either of unstinted praise or of utter condemnation?

Failure of artificial legs to satisfy our hopes results less from the imperfection of their mechanisms than from the extravagance of our expectations. People who do not expect a glass eye to see or a prosthetic hand to play the piccolo are disappointed when an artificial leg squeaks while dancing the polka. Man has never commanded clear appreciation of his means of locomotion. From time to time he has been ecstatic about the eye and the liver, the heart, the brain, the hand. Legs have been referred to most frequently as symbols of neighboring functions, so lightly have their own merits been regarded.

Why is the performance of the lower extremity so much less spectacular than that of the upper? Independence of the upper limb from obligation to the rest of the body allows it to indulge in ornamentation of movement, so impressive to the eye. The lower limb, sandwiched between the ground and the torso, must ever be responsive to the needs of the body as a whole. It cannot choose to support some parts of the body and not others or to walk with the body through only portions of each step. The intricacy of function of knee and ankle does not

¹ Associate Professor of Anatomy, College of Physicians and Surgeons, Columbia University; member, Lower-Extremity Technical Committee, ACAL, NRC.

exhibit itself in capricious movements but excels when it modulates countless disturbing factors so that no tremor mars the stark simplicity of normal locomotion.

No one can rightly expect an artificial limb to take over completely the functions of its predecessor unless it is endowed with an equivalent of muscular and nervous control. Difficult as it is to provide substitutes for bones and joints, such provision is simplicity itself compared with the incorporation within the prosthesis of its own control. Although considerable progress has been made in the field of decelerating mechanisms for lower-extremity prostheses, the leg amputee must still use his own resources when he needs to supply energy or to exercise discretion.

The contribution which the amputee makes to the over-all prosthetic result far exceeds that of acting as a model for exhibiting the achievements of inventors. It is he who must finish creation of the new locomotor mechanism by reshaping the pattern of his muscular activity and establishing alertness to new sensory cues. The success of the artificial leg depends on how thoroughly it becomes a part of the form and the function of the amputee after he has blended its metal, wood, and plastic with his muscle and perception. It is only appropriate that the new mechanism, having superseded the natural limb, should contribute to amputee gait that special accent which identifies the supernatural walk.

The complexity of human motion makes it inevitable that fundamental improvement in leg prostheses must come slowly, since it is based on factors so numerous that no one individual can comprehend them all. In addition to the profession of engineering, there is needed the cooperation of the physician, the physicist, the physiologist, the physiotherapist, the prosthetist, and the psychologist—to list them in alphabetical order—so that the patient may get the total care he deserves.

The problems which need attention are of different degrees of complexity and must be approached by different methods. Choice of materials, details of construction, and provisions for repair require less consideration of the over-all characteristics expected in the rehabilitated amputee than do problems of fit and socket shape. More general considerations must be weighed in projects concerned with alignment, basic design of mechanisms, and evaluation of performance. For these there should be a conscious choice of a realizable objective, the attainment of which requires integration of man and machine into a functional unit.

All of these are practical problems amenable to increasingly useful solutions year by year, provided we do not surrender to the impatience of those who must have the answer to the question of the century today and of the millennium tomorrow. It is necessary to preserve clear vision of long-term objectives, although some members of every team find the environment more familiar when details arise.

Had trial-and-error and serendipity been able to produce truly satisfactory lower limbs, we would not still be waiting for such. It was left for the National Academy of Sciences—National Research Council to initiate the development of artificial limbs on a modern basis by creating the Committee on Prosthetic Devices and, later, its successor, the Advisory Committee on Artificial Limbs. By carefully balancing the fundamental and the practical in their program, these Committees have laid a firm basis for some progress today, much more tomorrow.

This is the key to the future in lower-extremity prosthetics. Used wisely, it will allow us eventually to rescue the limb problem from limbo and to provide the amputee of the future with a fitting legacy.

The Objectives of the Lower-Extremity Prosthetics Program

HOWARD D. EBERHART, M.S.¹

MAN depends upon his legs to support the body and to move it from place to place as occasion warrants. Since mobility is nearly indispensable to most human activities, the loss of part or all of a leg—through accident, war, or disease—imposes serious limitations and has always made a replacement of some sort more or less of a necessity. Accordingly, artificial legs of one kind or another have been made and used since the most ancient times. As a result of the long-continued effort, leg prostheses have undergone progressive, if slow, development through the centuries, so that many lower-extremity amputees have in the past been successfully restored to something resembling a normal life. With the advent of industrial development, and of improved tools and materials with which to work, the nineteenth century marked the appearance of many new lower-extremity devices and of new techniques in the medical treatment of amputations.

Impetus provided by World Wars I and II gave rise to rapid advancement in all branches of technology and thus made possible a concerted attack on the problem of supplying the best possible artificial limbs. The term "lower-extremity prosthetics" has now come to mean the practice of rehabilitation of the leg amputee by providing him with an artificial limb that will restore lost functions to the greatest possible degree. But more than just the artificial leg is involved. The amputee himself is a

most important part of the end-product, and amputees, like other people, are individuals with widely differing characteristics and abilities. Of course surgical procedures should be designed to secure a painless stump and to retain maximum function, and it would seem that the artificial leg, when properly fitted, should duplicate as closely as possible the normal activity of the lost part. Moreover, physical conditioning and gait training are both important phases of the whole rehabilitation process.

This concept of lower-extremity prosthetics has developed during the years since the start in 1945 of the program of the Advisory Committee on Artificial Limbs, National Research Council. Initially, the primary objective was to develop improved devices, it being considered as obvious that, if a better prosthetic knee or ankle or foot could be devised, the amputee would benefit. Attempts to produce such items, however, made necessary the determination of functional requirements and thus immediately revealed the lack of necessary fundamental information. Basic research into the complicated phenomenon we call "locomotion" was therefore carried on simultaneously with the development of devices.² These investigations indicated a need for the application of basic mechanical principles to fitting and alignment of artificial legs. A three-

¹ Professor of Civil Engineering, University of California, Berkeley; member, Advisory Committee on Artificial Limbs, National Research Council; chairman, Lower-Extremity Technical Committee, ACAL, NRC.

² A more logical and systematic approach, had there been sufficient time, might have been to postpone device development until the results of the basic work became available. But the urgency of amputee demands at the end of World War II made such an approach less desirable than the one adopted.